Technical Data

The technical data of the Sorvall ST 16 / 16R is listed in the following table.

Table 1-1. Technical Data Sorvall ST 16 / 16R

Feature	Value			
Environmental conditions	- use in interior spaces			
	- altitudes of up to 2,000 m above sea level			
	- max. relative humidity 80% up to 31 °C; decreasing linearly up to 50% relative humidity at 40 °C.			
permissible ambient temperature	+2° C to +35° C		+2° C to +35° C	
Overvoltage category	II			
Pollution degree	2		2	
Heat dissipation	refrigerated		ventilated	
	230V	120V	230V	120V
	4778 BTU/h	4096 BTU/h	3447 BTU/h	3901 BTU/h
IP	20		20	
running time	unlimited		unlimited	
maximum speed n _{max}	15200 rpm		15200 rpm	
minimum speed n _{max}	300 rpm		300 rpm	
maximum RCF value at n _{max}	25830xg		25830xg	
maximum kinetic energy	<62,5 kJ		<51,7 kJ	
noise level at maximum speed	<63 dB(A)		<63 dB(A)	
temperature setting range	-10 °C to +40 °	-10 °C to +40 °C		
Dimensions	refrigerated		ventilated	
Height	360 mm		360 mm	
Height lid open	870 mm		870 mm	
Width	623 mm		440 mm	
Depth	605 mm 310 mm		605 mm	
Table top height			310 mm	
weight without rotor	91,5 kg		57,5 kg	

Thermo Scientific ST16 / 16R 1-3

Directives, standards and guidelines

Table 1-2. Directives, standards and guidelines

Tension / Frequency		Produced and inspected according to the following standards and guidelines
230V 50/60Hz	2006/95/EC Low Voltage Directive: 2006/42/EC & 98/37/EC Machine Directive: 2004/108/EC EMC Directive 98/79/EC In-vitro-Diagnostika (IvD)	EN 61010-1, 2 nd Edition EN 61010-2-020, 2 nd Edition EN 61010-2-101 EN 61326-1 EN 61326-2-6 EN 55011B EN 61000-6-2 EN ISO 13485
230V 60Hz 120V 60Hz 100V 60Hz	The centrifuges are produced and inspected according to the following standards and guidelines	UL 61010-1, 2 nd Edition CAN/CSA-C22.2 No. 61010-1, 2 nd Edition IEC 61010-2-20, 2 nd Edition (Pollution degree 2, Overvoltage category II) IEC 61010-2-101 Emitted interference FCC Part 15 CLASS A NOTE: This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense.
100V 50Hz		IEC 61010-1 2 nd Edition IEC 61010-2-020 2 nd Edition IEC 61010-2-101 EN 61326-1 EN 61326-2-6 EN 55011A EN 61000-6-2 EN ISO 13485

1-4 ST16 / 16R Thermo Scientific