

5. Fault Diagnosis

Symptom	Possible Cause	Action Required
1. Unit does not operate	a. Unit not switched on b. Unit not plugged into power supply c. Fuses blown d. Power supply failure	a. Switch on b. Plug in, turn on c. Replace fuses per 8.2 d. Check that other electrical appliances on the same circuit are working
2. Bath temperature does not rise when expected	a. Actual temperature is higher than set temperature b. Temperature control sensor is not positioned in the water c. Temperature control circuit fault	a. Check set temperature b. Position sensor in the water. c. Have unit checked by competent person
3. Temperature continues to rise when not expected	a. Actual bath temperature is lower than set temperature b. Temperature control circuit fault	a. Check set temperature b. Have unit checked by competent person
4. Lamp not illuminated	a. Lamp not switched on b. Lamp failure	a. Press lamp switch b. Have unit checked by competent person

6. Accessories

Part Number	Description
904-0013	Glass Dish

7. Technical Specifications

This equipment is intended for indoor use and will meet its performance figures within the ambient temperature range of 20°C to 30°C, with maximum relative humidity of 80% (non-condensing). Installation Category II (transient voltages). Pollution Degree 2 in accordance with IEC 664. Suitable for operation at altitudes of up to 6500 feet.

Specifications:

Temperature Set Range:	(Ambient +5°C) to 65°C
Operating Temperature Range:	20°C to 30°C
Temperature accuracy:	± 0.3°C at 55°C
Stability:	± 0.2°C
Supply Voltage Range:	
145702 and 145951	115V ± 10%, 60 Hz
145702-2 and 145951-2	230V ± 10%, 50/60Hz
Power Rating:	272W
Heating Rate:	Ambient to 55°C within 45 minutes