

1 PRODUCT DESCRIPTION

1.2.1.3 Installation

Follow these steps:

- Check whether power supply corresponds with the one named on the manufacturer's rating label which is mounted on the rear panel.
- The line voltage circuit breaker is max. 10 A (type K) slow release for commonly used instruments.
- In case of emergency, there must be an emergency switch off installed outside the room in order to disconnect the power supply of the unit.

(•Remove the transport spacer blocks from the motor shaft (see chapter 2.2.2).)

The outlet for the power cord must be easy to reach and to disconnect!

1.3 Technical Data

Manufacturer	HERMLE Labortechnik GmbH	
Type / Model	Z 206 A	
Dimensions		
Width	28 cm	
Depth	37 cm	
Height	26 cm	
Weight	15 kg	
Noise level (max.)	60 +2.0 dB (A)	
Max. speed	6000 rpm	
Max. volume	6 x 50 ml	
Max. RCF	4185 x g	
Admissible density	1.2 kg/dm ³	
Admissible kinetic energy	1694 Nm	
Electrical connection AC	230 V / 50 Hz 1 ph	120 V / 60 Hz 1 ph
Current	0.55 A	1.1 A
Connected load	100 Watt	100 Watt
Interference suppression	VDE 0871, Funkentstörgrad B	
Test obligations		
To be filled in by purchaser:		
Inventory-No.:	_____	
Check-No.:	_____	
Location:	_____	
Maintenance contract:	_____	
Your service department	Labnet International PO Box 841, Woodbridge, NJ Ph: 732-417-0700	
Your agent	_____	